

Survey: KS4/16  
month: May 2016  
Participant-No: 0004965  
date of issue: 23.05.16



Referenzinstitut für Bioanalytik

diaglobal GmbH  
Dr. Rainer Schlegel  
Haus 41  
Köpenicker Str. 325  
12555 Berlin



Survey directors

Prof. Dr. C. Knabbe  
Prof. Dr. Dr. K.P. Kohse  
Prof. Dr. M. Neumaier

Head of RfB

Dr. R. Kruse  
Dr. W. J. Geilenkeuser

Bonn, 13. June 2016

## Certificate

We confirm that you have participated in the survey for clinical chemical analytes in serum.

You have met the requirements of the survey for the following analytes:

bilirubin, total	(2)	cholesterol	(1)	glucose	(3)
urea	(1)	creatinine	(1)	lactate	(2)
triglycerides	(2)	calcium	(3)	magnesium	(4)
CK	(1)				

This certificate is valid until the end of November 2016.

*Rainer Schlegel* *Rainer* *W.J. Geilenkeuser*

The number in parentheses characterizes the analytical method used.

The assignment of the number to the respective method and/or the respective instrument is to be taken from the total evaluation.

## Listing and Evaluation of all your results

### Explanations

#### Certificate

A certificate is issued (given) for an analyte only if the basis for an evaluation of the accuracy is given by the guidelines of the German Medical Association and/or if an evaluation is possible in analogy to these guidelines (see comments on the evaluation) and, if both results for an analyte are within the given acceptance limits. (marked as '+' below C)

#### Certificate of participation

In the participation certificate all analytes which are included in the list on this page are listed.  
 If all analytes are listed on the certificate no participation certificate is printed.

Legend:	C=Certification, M=No of method, R=your result, D=difference (R-T) Dmax= maximum allowable amount of difference in measurement; partly given by annex 1 of the guideline of the BÄK (Dt. Ärzteblatt 111, Heft 38, 19.9.2014). T = target value, either reference method value or assigned value. LL   UL = lower resp. upper limit	Certification:	+ = fulfilled ( quotient   D/Dmax   <= 1.0 ) - = not fulfilled (quotient   D/Dmax   > 1.0 ) ± = certification cancelled because of technical and/or analytical reasons
---------	---	----------------	---

	C	M	R	D/Dmax	T	LL	UL					
bilirubin, total [mg/dl]	+	2	A 4.43 B 1.99	0.31 0.56	4.14 1.77	3.22 1.38	5.06 2.16					
cholesterol [mg/dl]	+	1	A 182 B 174	-0.32 0	190 174	165 151	215 197					
glucose [mg/dl]	+	3	A 196 B 73.3	-0.38 -0.19	208 75.5	176 64.1	240 86.9					
urea [mg/dl]	+	1	A 61.0 B 190	0.09 0.02	59.9 189	47.9 151	71.9 227					
creatininine [mg/dl]	+	1	A 2.64 B 4.29	-0.16 -0.24	2.73 4.51	2.18 3.60	3.28 5.42					
lactate [mmol/l]	+	2	A 4.98 B 2.09	0.08 0.24	4.90 2.00	4.00 1.63	5.79 2.37					
triglycerides [mg/dl]	+	2	A 143 B 96.2	-0.29 0.03	150 95.6	126 80.3	174 111					
calcium [mmol/l]	+	3	A 3.03 B 3.59	0 -0.03	3.03 3.60	2.72 3.24	3.34 3.96					
magnesium [mmol/l]	+	4	A 1.86 B 0.770	-0.24 0.46	1.93 0.720	1.64 0.612	2.22 0.828					
CK [U/l]	+	1	A 545 B 537	-0.24 -0.31	573 586	458 427	688 745					