

Survey: KS4/18
month: May 2018
Participant-No: 0004965
date of issue: 19.05.18



Referenzinstitut für Bioanalytik

diaglobal GmbH
Dr. Rainer Schlegel
Haus 41
Köpenicker Str. 325
12555 Berlin



Survey directors

Prof. Dr. C. Knabbe
Prof. Dr. Dr. K.P. Kohse
Prof. Dr. M. Neumaier

Head of RfB

Dr. W. J. Geilenkeuser
Dr. A. Kessler

Bonn, 13. June 2018

Certificate

We confirm that you have participated in the survey for clinical chemical analytes in serum.

You have met the requirements of the survey for the following analytes:

bilirubin, total	(2)	cholesterol	(1)	glucose	(3)
urea	(1)	creatinine	(1)	lactate	(2)
calcium	(3)	magnesium	(4)	CK	(1)
AST	(1)	ALT	(1)		

This certificate is valid until the end of November 2018.

The number in parentheses characterizes the analytical method used.

The assignment of the number to the respective method and/or the respective instrument is to be taken from the total evaluation.

http://www.rfb.bio/certificates?doc=0004965&v_h=KS4/18&version=3&uid=A48163D0818B4F8C98E77F0D0287012

Survey: KS4/18
 month: May 2018
 Participant-No: 0004965



Listing and Evaluation of all your results

Explanations

Certificate

A certificate is issued (given) for an analyte only if the basis for an evaluation of the accuracy is given by the guidelines of the German Medical Association and/or if an evaluation is possible in analogy to these guidelines (see comments on the evaluation) and, if both results for an analyte are within the given acceptance limits. (marked as '+' below C)

Certificate of participation

In the participation certificate all analytes which are included in the list on this page are listed.
 If all analytes are listed on the certificate no participation certificate is printed.

Legend:	C=Certification, M=No of method, R=your result, D=difference (R-T) Dmax= maximum allowable amount of difference in measurement, partly given by annex 1 of the guideline of the BÄK (Dt. Ärzteblatt 111, Heft 38, 19.9.2014). T = target value, either reference method value or assigned value, LL UL = lower resp. upper limit	Certification: + = fulfilled (quotient D/Dmax <= 1.0) - = not fulfilled (quotient D/Dmax > 1.0) ± = certification cancelled because of technical and/or analytical reasons
---------	--	--

	C	M	R	D/Dmax	T	LL	UL								
bilirubin, total [mg/dl]	+	2	A	1.85	0.56	1.64	1.27	2.01							
			B	3.53	0.46	3.20	2.49	3.91							
cholesterol [mg/dl]	+	1	A	147	-0.30	153	133	173							
			B	131	-0.17	134	116	152							
glucose [mg/dl]	+	3	A	75.6	0.00	75.5	64.1	86.9							
			B	149	-0.42	159	135	183							
urea [mg/dl]	+	1	A	193	-0.03	194	155	233							
			B	142	-0.35	153	122	184							
creatinine [mg/dl]	+	1	A	4.38	-0.04	4.42	3.53	5.31							
			B	5.49	-0.16	5.67	4.53	6.81							
lactate [mmol/l]	+	2	A	2.04	0.16	1.98	1.60	2.35							
			B	3.46	0.25	3.31	2.70	3.91							
triglycerides [mg/dl]	-	2	A	66.6	-1.17	82.1	68.9	95.3							
			B	97.2	-0.78	111	93.2	129							
calcium [mmol/l]	+	3	A	3.54	0.20	3.47	3.12	3.82							
			B	2.58	0.26	2.51	2.25	2.77							
magnesium [mmol/l]	+	4	A	0.810	0.78	0.725	0.616	0.834							
			B	1.48	0.09	1.46	1.24	1.68							
CK [U/l]	+	1	A	448	-0.75	527	421	633							
			B	320	-0.62	365	292	438							
AST [U/l]	+	1	A	99.4	-0.08	101	79.7	123							
			B	246	0.07	242	191	293							
ALT [U/l]	+	1	A	242	-0.29	258	203	313							
			B	212	-0.37	230	181	279							