

Survey: **BI2/26**
month: **May 2026**
Participant-No: **0004965**
valid from: **02.05.26**



page 1 of 3

diaglobal GmbH
Dr. Rainer Schlegel
Haus 41
Köpenicker Str. 325
12555 Berlin



Bonn, 8. May 2026

Certificate

We confirm that you have participated in the survey for neonatal bilirubin in serum.
This certificate is according to the current version of the RiliBÄK valid until the end of November 2026.

You have met the requirements of the survey for the following analytes:

Total bilirubin (2)

Handwritten signature of Prof. Dr. Dr. K. P. Kohse in blue ink.

Prof. Dr. Dr. K. P. Kohse
EQA scheme director

Handwritten signature of Dr. Anja Kessler in blue ink.

Dr. Anja Kessler
Head of Reference Institute


Handwritten signature of Dr. Marika Enders in blue ink.

Dr. Marika Enders
EQAS-Board

The number in parentheses characterizes the analytical method used.
The assignment of the number to the respective method and/or the respective instrument is to be taken from the total evaluation.

This certificate is valid in conjunction with the final report dated 08.05.26.
This is available for download in the RfB-Online system.

Analyte **Total bilirubin**
 Method all methods



No of participants	435		
sample/unit	A	mg/dl	B
mean	2.35		12.4
standard deviation	0.229		1.03
coefficient of variation	9.75		8.32

Sample A (RMW = 2.37 mg/dl)

M	Kit	N	Min	16.P	50.P	84.P	Max	
Alle		435	0.700	2.14	2.36	2.60	3.10	1,4 2,1 2,8 3,5
1	30	7	2.10		2.13		2.21	
1	328	4	2.50		2.55		2.57	
2	4	40	1.56	2.29	2.40	2.50	2.60	
2	30	142	1.56	2.08	2.17	2.25	2.51	
2	38	19	2.20	2.29	2.36	2.42	2.55	
2	43	4	2.00		2.07		2.10	
3	99	3	1.70		2.70		2.92	
3	115	84	2.10	2.44	2.60	2.70	3.10	
6	43	18	2.08	2.40	2.46	2.56	2.70	
7	32	6	2.30		2.60		2.60	
7	61	89	0.700	2.30	2.40	2.51	2.90	

Sample B (RMW = 12.3 mg/dl)

M	Kit	N	Min	16.P	50.P	84.P	Max	
Alle		435	0.540	11.4	12.6	13.3	24.3	6 9 12 15 18
1	30	7	11.0		11.4		12.1	
1	328	4	13.0		13.8		14.3	
2	4	40	10.3	12.6	13.0	13.4	13.9	
2	30	142	0.540	11.2	11.6	11.9	13.2	
2	38	19	11.0	11.6	11.8	12.0	12.9	
2	43	4	10.9		11.3		11.7	
3	99	3	9.69		12.0		13.7	
3	115	84	8.80	12.9	13.2	13.6	24.3	
6	43	18	9.50	12.8	13.6	14.4	15.2	
7	32	6	12.7		13.4		13.9	
7	61	89	3.30	12.4	12.8	13.2	14.2	

Other kits (number):
 1-04(1), 1-43(1), 1-63(1), 1-115(1), 2-12(2), > 2-83(1)<, 2-99(1), 3-01(1), 3-02(1), 3-07(1), 3-15(1), 3-83(1),
 3-108(1), 5-166(1), 6-89(2), 7-63(1), 7-99(1),